Alinity m HSV 1 & 2 / VZV – a new qualitative multi-plex PCR test

Robert Ehret, Malik Prentice, Özlem Yildirim, Martin Obermeier

Medical Center for Infectious Diseases, Berlin (MIB)

Contact: Ehret@mvz-mib.de

BACKGROUND

HSV-1, HSV-2 and VZV cause self-resolving skin or mucosal diseases and concurrently establish persistent latent infections in sensory ganglia neurons. HSV-1 primarily affects the oral area and is mainly transmitted through oral-to-oral contact; however, the virus can also cause genital herpes. HSV-1, HSV-2 and VZV often cause phenotypically indistinguishable skin blisters or lesions. VZV lesions typically localized in specific anatomical areas (dermatomes) can also spread to other areas, including the genital region.

Tab. 1: EQA swab tested in triplicates with Allplex and Alinity

EQA sample	Allplex HSV-1	Allplex HSV-2	•	Alinity HSV-1	-	Alinity VZV
sampre						
1	23,98	3 20,78	8 23,94	18,58	3 16,13	20,28
2	24,08	3 20 <i>,</i> 93	1 23,75	18,83	8 16,28	20,30
3	24,00) 20,72	2 23,69	18,69	9 16,17	20,15
Allplex TP	28,00) 28,10	28,20	_	-	_

We evaluated the performance of Alinity m HSV 1&2 / VZV, a new fully automated qualitive multiplex real-time PCR developed for simultaneous detection and differentiation of the three pathogens in clinician-collected swab specimens from cutaneous and mucocutaneous lesions.


METHODS

A HSV-1&2/VZV/Syphilis Positive Swab EQA sample (Microbix Biosystems Inc.) resuspended in 3 ml 0.9% NaCl was tested in triplicate with Alinity m HSV 1&2/VZV and Allplex Genital ulcer assays to compare their response to the three pathogens targeted by the Alinity assay. Assay linearity and precision were assessed by testing 10 replicates of a dilutional series prepared from HSV-1/2 and VZV EQA samples (2.000.000 cp/mL of inactivated virus each; Exact Diagnostics Verification Panel, BioRad) with concentrations ranging from 5000 to 20 cp/mL for each virus and mixture of the 3 viruses.

A total of 360 residual archived (-20°C) clinical samples (lesion swabs [N=206], urine [N=21] and pooled urine/swab samples [N=81] collected in Aptima Specimen Collection Kits (Hologic), pre-selected based on previous results established during laboratory routine testing with the Allplex Genital ulcer Assay (Seegene), detecting HSV-1/2, VZV, *T. pallidum*, *H. ducreyi*, LGV and CMV and the Allplex STI Essential Assay (Seegene), detecting *C. trachomatis*, *N. gonorrhoeae*, *M. genitalium*, *M. hominis*, *T. vaginalis*, *U. urealyticum* and *U. parvum*. All specimens were parallel-tested with Alinity m and Allplex Genital ulcer assays; 42 samples with insufficient volume were diluted with Aptima Collection fluid (maximum factor 1:2) prior to testing.

Tab. 2: Dilution series with Exact Diagnostics Verification Panel

HSV-1	Allplex	Allplex	Allplex	Allplex	Alinity	Alinity	Alinity	Alinity
(cp/ml)	Det Rate	Mean	SD	%CV	Det Rate	Mean	SD	%CV
5000	100%	27,48	0,72	2,62%	100%	25,04	0,65	2,60%
500	100%	31,22	1,79	5,73%	100%	28,14	0,41	1,46%
100	50%	33,04	1,66	5,32%	50%	30,52	0,93	3,05%
50	20%	32,86	0,49	1,48%	20%	30,73	0,31	1,01%
20	20%	33,49	0,53	1,61%	0%			
HSV-2	Allplex	Allplex	Allplex	Allplex	Alinity	Alinity		Alinity
(cp/ml)	Det Rate	Mean	SD	%CV	Det Rate	Mean	Alinity SD	%CV
5000	100%	32,92	0,65	1,97%	100%	26,43	0,24	0,91%
500	60%	36,7	0,84	2,29%	95%	30,09	0,45	1,50%
100	25%	37,7	0,6	1,63%	50%	32,18	0,47	1,46%
50	5%	37,6			35%	32,3	0,26	0,80%
20	5%	37,4			10%	32,94	0,01	0,03%
VZV	Allplex	Allplex	Allplex	Allplex	Alinity	Alinity	Alinity	Alinity
(cp/ml)	Det Rate	Mean	SD	%CV	Det Rate5	Mean2	SD3	%CV4
5000	95%	32,73	0,99	3,02%	100%	30,2	•	1,26%
2500	40%	33,39	0,26	0,78%	100%	31,53	0,3	0,95%
2000	80%	34,32	0,46	1,38%	40%	32,59	0,02	0,06%
1500	0%				30%	32,62	0,31	0,95%
1000	0%				40%	32,4	0,02	0,06%
500	0%				5%	32,75		
100	0%				0%			




RESULTS

Both assays detected the Herpes viruses in the EQA positive swabs at high concentrations with the lower Ct-values for Alinity (3-5 cycles, see Tab. 1). Both assays demonstrated high dilutional sensitivity and linearity, with slightly lower sensitivity for VZV. No compensatory effects were seen in triple-positive samples compared to single-positive samples. The mean Ct-values of the ten measurements of the 5000 cp/mL dilution were identical in the single virus measurement and the mixture of all three viruses. Overall, precision was higher with the Alinity assay (see Tab. 2).

All samples pretest-negative for HSV-1/2 and VZV [N=100] tested negative with the Alinity assay. No cross-reactivity for non-targeted sexually-transmitted pathogens present in 144 clinical samples (most often *U. urealyticum* and *C. trachomatis* and up to four different additional microorganisms) in the absence or presence of HSV-1/2 or VZV was observed with the Alinity test. For HSV-1, 101 of 109 pretest-positive samples were confirmed by at least one subsequent test. For HSV-2 127 of 128, and all VZV were confirmed. Overall-agreement between Alinity and Allplex for qualitative detection of HSV-1, HSV-2 and VZV in clinical samples was 95.4% (104/109), 98.4% (126/128), and 100% (30/30), respectively (see Figures 1-3). Alinity identified 4 infections missed by Allplex, while 3 infections were only detected by Allplex.

CONCLUSIONS

The Alinity m HSV 1 & 2 / VZV assay demonstrated excellent efficacy in detecting the Herpes viruses, exhibiting no discernible competitive effects in samples with co-infections. The test showed good linearity and higher sensitivity for VZV and HSV-2 compared to the Allplex Genital Ulcer Assay. The random-access capability of the Alinity platform is a valuable tool in clinical routine.



ACKNOWLEDGEMENTS: The study was funded by Abbott GmbH, Wiesbaden, Germany.



Medical Center for Infectious Diseases, Berlin

